



**Technical Evaluation Report for the Procurement of Infant Warmer on CIF basis 2018-19
under IRMNCH & Nutrition Program, Punjab dated 03.05.2019**



Sr.	Firm Name	Quoted Model	Mfg. By	Technical Evaluation													Sample Specification Evaluation	Result	Remarks
				1	2	3	4	5	6	7	8	9	10	11	12	13	1		
				Legal registered entity	Active Tax Payer (NTN & GST) as per clause 3 ITB	One person One Bid	Manufacturer/ Sole Distributor	Past performance	Balance Sheet, Financial Statement & Annual Turnover	Complete item with accessories and optional	Certificate from the Manufacturer regarding after Sales Services and Installation	Testing/Calibration equipment	Affidavit from Bidder	Bid Validity	Delivery Period	Compliance of Warranty as per tender	Specification Compliance		
1	Medi Urge (Pvt) Ltd	Baby Guard W-1140	Dixon, Couvenstr.6, 40211, Dusseldorf, Germany	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	Yes	Yes	Yes	Yes	Yes	Yes	Approved	Responsive	
2	Vertex Medical	Babytherm 8004	Dragerwerk AG&Co. KGaA, Germany	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	Yes	Yes	Yes	Yes	Yes	Yes	Approved	Responsive	
3	Alam Medix	BabyLife BLR-2100	Medicor Elektronika Zrt. Hungary	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	Yes	Yes	Yes	Yes	Yes	Yes	Not Approved	Non-Responsive	Contradiction between brochure and website. Heating power and Manual heat output not as per tender requirements.
4	Medequips SMC	Sunflower Warmer	Atom Medical International Inc. Hongo, Bunkyo, Tokyo, Japan	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	Yes	Yes	Yes	Yes	Yes	Yes	Approved	Responsive	
5	Radiant Medical	Infant Warmer SM-401	Natus Medical Incorporated, 1501 Industrial Road, San Carlos, USA	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	Yes	Yes	Yes	Yes	Yes	Yes	Not Approved	Non-Responsive	Attached documents shows Ireland base company whereas claimed USA. No quality certificate for quoted model.
6	Shirazi Trading	Lullaby Warmer	General Electric Company	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	No	Yes	Yes	Yes	Yes	Yes	Approved	Non-Responsive	Quoted Product does not have certificate of quality i.e. FDA/CE/PPMHLW. Invalid Certificate from the Manufacturer regarding after Sales Services and Installation attached.
7	Quintex Medicals	Infant Radiant Warmer BN-100A	Fazzini SRL, S.S. Padana Superiore, Vimodrone (MI), Italy	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	Yes	No	Yes	Yes	Yes	Yes	Not Approved	Non-Responsive	The firm did not attach list of testing/calibration equipment. BN-100 is also available with same specification on the website of Dison China. Manual heat output control not available. Material of heating source not mentioned.
8	Delta Plus	Infant Warmer-A4051	Advanced Instrumentations, 6800 N.W. 77 Court, Miami, FL 33166 U.S.A	Yes	Yes	Yes	Sole Distributor	Yes	Yes	As per sample/ Brochure	Yes	Yes	Yes	Yes	Yes	Yes	Not Approved	Non-Responsive	As per USFDA Establishment Registration & Device Listing, the company which purports to be the manufacturer of the quoted product is enlisted as Repackager/Relabeler

	Sr.	Parameters							
Technical Evaluation	1	Legal registered entity							
	2	Active Tax Payer (NTN & GST) as per clause 3 ITB							
	3	One person One Bid							
	4	Manufacturer /Sole Distributor							
	5	Past performance							
	6	Balance Sheet, Financial Statement & Annual Turnover							
	7	Complete item with accessories and optional							
	8	Certificate from the Manufacturer regarding after Sales Services and Installation							
	9	Testing/Calibration equipment							
	10	Affidavit from Bidder							
	11	Bid Validity							
	12	Delivery Period							
	13	Compliance of Warranty as per tender							